
Power-Stain™ 1.0 Double Stain Kit II (Poly AP for Mouse + Poly HRP for Rabbit)

<u>Cat No.</u>	<u>Quantity</u>
52-0024	15 mL

Intended Use For In Vitro Diagnostic Use.

This kit is intended for use with Mouse and Rabbit Primary Antibodies and other ancillary reagents supplied by user for qualitative detection of targeted proteins (antigens) using immunohistochemistry (IHC) methodology by light microscopy on routine formalin-fixed, paraffin-embedded (FFPE) tissue section.

Interpretation of any positive or negative staining shall be supported by implementation of a proper control, and must be made within the context of the patient's clinical history and other diagnostic test by a qualified pathologist.

Summary And Explanation

This kit is a non-biotin system and utilizes a Poly AP (alkaline phosphatase) and Poly HRP (horseradish peroxidase) conjugates to locate where the mouse or rabbit primary antibody is bound to the target antigens. The complexes formed between Poly AP conjugate with mouse primary antibody and Poly HRP conjugate with rabbit primary antibody are observed through the use of two substrate-chromogen solutions, which when added sequentially, result in two colored precipitates at the antigen locations. The staining location and pattern is easily observable by light microscopy.

Reagents Supplied

One bottle of ready-to-use **Poly AP Conjugate for Mouse + Poly HRP for Rabbit** in an enzyme conjugate buffer containing stabilizing proteins and anti-microbial agents.

Storage

Store at 2-8°C. Do not freeze.

All performance claims are void after the kit expiration date.

Materials Required But Not Supplied

Mouse and Rabbit Primary Antibodies (Genemed offers prediluted and concentrate Primary Antibodies)
Primary Antibody Diluent (Cat No. 10-0001)

Reagent Control (Non-immune Mouse IgG Cat No. 60-0045-7 and Non-Immune Rabbit IgG Cat No. 60-0060-7)

Positive and Negative Control Specimens

Microscope Slides, Positively Charged

Xylene

Ethanol

Endogenous Peroxidase Blocking Solution - 3% Hydrogen Peroxide (Cat No. 10-0056)

Wash Buffer - 10 mM Phosphate Buffer Saline, pH 7.4; optional with 0.05% Tween 20

Substrate-Chromogen Solution for HRP Conjugate (e.g. Cat No. 10-0006 DAB Substrate Kit; Cat No. 10-0048 Sensitive DAB Substrate Kit; Cat No. 10-0005, 10-0047 AEC Substrate Single Solution)

Substrate-Chromogen Solution for AP Conjugate (e.g. Cat No. 10-0008 Liquid Fast Red Substrate Kit; 10-0007 BCIP-NBT Substrate Kit)

Hematoxylin (Cat No. 10-0027, 10-0049)

Antigen retrieval reagents (e.g. Cat No. 10-0022 Citrate Buffer pH 6.0 1X; Cat No. 10-0020 Citrate Buffer pH 6.0 20X; Cat No. 10-0021 Tris Buffer pH 9 20X; Cat No. 10-0023 Tris Buffer pH 9 1X; Cat No. 10-0046 Tris EDTA Buffer pH 9 1X; Cat No. 10-0037 Tris EDTA Buffer pH 9 20X; Cat No. 10-0024 Proteinase K; Cat No. 10-0025 Trypsin; Cat No. 10-0050 Ficin)

Precautions

For professional users only.

Sodium Azide inhibits peroxidase activity. Use caution when handling Poly HRP and Poly AP Conjugate to prevent any contamination with other reagents containing sodium azide.

Proper handling of this product as with any product derived from biological sources should be used according to local and applicable regulations.



Procedural Notes The directions accompanying this kit provide step by step instructions for optimal staining. Any change in procedure or incubation times may give erroneous staining results. For optimal results, do not substitute the reagent provided in the kit.

Reagent shall be equilibrated to room temperature readily before usage. All incubations should be performed at room temperature in a humid environment.

Do not allow the tissue section to dry out at any point in the staining procedure. The reagent is for single use.

Preliminary Preparation Of Slides

Routine de-paraffinization and rehydration of tissue section.

Antigen retrieval as required by the primary antibody.

Control Slides

Three types of control slides are necessary for proper interpretation.

Positive Tissue Control – A tissue containing the desired antigen.

Negative Tissue – A tissue that does not contain the desired antigen.

Reagent Control – A slide to be treated with a homologous non-immune immunoglobulin.

(Cat No. 60-0045-7 or Cat No. 60-0060-7)

Staining Protocol

Step 1: Endogenous Peroxidase Blocking

- a) Submerge slides in Peroxidase Blocking Solution for 10 minutes.
- b) Wash slides with Wash Buffer to remove excess Peroxidase Blocking Solution.
- c) Tap off excess liquid and carefully wipe around tissue.

Step 2: Primary Antibody Incubation

- a) Prepare Primary Antibody Cocktail (mouse and rabbit) to optimum concentration. If necessary, dilute with Primary Antibody Diluent.
- b) Add 2 drops (100 μ L) or as much as needed of Primary Antibody Cocktail to completely cover each tissue.
- c) Incubate for 30-60 minutes at room temperature.
- d) Rinse 3 times with Wash Buffer for 2 minutes each.
- e) Tap off excess liquid and carefully wipe around tissue.

Step 3: Poly AP + Poly HRP Conjugate Incubation

- a) Add 2 drops (100 μ L) or as much as needed of Enzyme Conjugate to completely cover each tissue.
- b) Incubate for 15 \pm 1 minutes.
- c) Rinse 3 times with Wash Buffer for 2 minutes each.
- d) Tap off excess liquid and carefully wipe around tissue.

Step 4: Substrate/Chromogen

- a) Prepare chromogen solutions according to product's specification sheet.
Chromogens recommended: DAB substrate solution for HRP Conjugate and Fast Red substrate solution for AP Conjugate.
- b) Add 100 μ L of substrate solution for HRP conjugate (e.g. DAB) or as much as needed to completely cover each tissue. Refer to product's specification sheet for incubation time.
- c) Rinse 3 times with Wash Buffer for 2 minutes each to remove excess substrate solution.
- d) Add 100 μ L of substrate for AP conjugate (e.g. Fast Red) or as much as needed to completely cover each tissue. Refer to product's specification sheet for incubation time.
- e) Rinse slides with tap water to remove excess substrate solution.
- f) Proceed with normal counterstaining and mounting protocol.

Step 5: Counterstaining

- a) Counterstain according to manufacturer's instruction.

Note: Some Chromogens, e.g. Fast Red, forms an end-product which is soluble in organic compounds. Therefore, it is necessary to use a nonalcoholic counterstain, such as Mayer's or Gill's Hamatoxylin, and an aqueous based mounting medium. Do not use mounting media



containing organic solvent.

Step 6: Mounting

- a) Mount and coverslip the specimen with appropriate mounting.

Interpretation Of Staining Results

Step 1: Review Positive and Negative Controls. Do not proceed to next step if the staining intensity does not meet requirements.

Step 2: Score the tested specimens.

Troubleshooting

	Positive Control Tissue	Negative Control Tissue	Reagent Control	Test Tissue	Analysis of Result
1	+	--	--	+	Specimen contains the antigen
2	+	--	--	--	Specimen does not contain the antigen

	Positive Control Tissue	Negative Control Tissue	Reagent Control	Test Tissue	Analysis of Result
1	--	--	--	--	No staining
2	Weak +	--	--	+/-	Weak staining
3	+	+	+	+	High background staining

Possible causes and suggested action for: No staining on any slide

1. Reagents not used in correct order.
→ Repeat procedure following Staining Protocol Instructions.
2. Substrate-Chromogen reagent(s) not prepared properly.
→ Prepare a fresh Substrate-Chromogen solution following product's instructions.
3. Primary antibody incubation steps were omitted or dilution was incorrect or wrong antibody was used.
→ Repeat procedure following Staining Protocol Instructions using incubation times specified.
→ Repeat procedure using correct dilution for primary antibody or correct primary antibody.
4. Wrong Pretreatment.
→ Repeat procedure using correct pretreatment.

Possible cause and suggested action for: Weak staining on all slides

1. Substrate-Chromogen reagent(s) has expired.
→ Prepare a fresh Substrate-Chromogen solution following product's instructions.
2. Incubation times were not long enough.
→ Repeat procedure following Staining Protocol Instructions using incubation times specified.
3. Specimen retained too much liquid after rinsing steps.
→ Tap off excess liquid and carefully wipe around specimen after rinsing steps.
4. Enzyme Conjugate exposed to sodium azide.
→ Use buffer without sodium azide, or check if reagent is contaminated with sodium azide during use or aliquot/pipetting.
5. Primary antibody dilution was incorrect.
→ Repeat procedure following Staining Protocol Instructions using incubation times specified.
→ Repeat procedure using correct dilution for primary antibody.
6. Insufficient Pretreatment.
→ Repeat procedure using correct pretreatment.

Possible cause and suggested action for: High background staining on all slides

1. Specimens contain high endogenous peroxidase activity.
→ Check preparation of Peroxidase Solution and verify timing of specimens submerged in solution.
2. Inadequate rinsing of slides.
→ Use freshly prepared buffer solutions. Follow rinsing instructions specified.
3. De-paraffinization not complete.
→ Use fresh xylene. Check slides are de-paraffinized before rehydration step.
4. Over-reaction of substrate.
→ Do not incubate substrate longer than specified in procedure.
5. Specimens dry out during staining procedure.
→ Incubate in humid environment. Wipe fewer than 10 slides at a time before adding next solution.
6. Wrong Pretreatment.
→ Repeat procedure using correct pretreatment.

Symbols

				
Catalog No.	Batch No.	In Vitro Diagnostic Use	Temperature Range	Use By